



TEST CERTIFICATE

ISSUED BY SIRA TEST & CERTIFICATION LIMITED

TESTS FOR THE INGRESS PROTECTION OF A QRAE3 PUMPED GAS SENSOR

Customer: RAE Systems
Kirstinehoj 23A
2770 Kastrup
Denmark

Type identification: QRAE3 Pumped Gas Sensor Main Enclosure

Standard: BS EN 60529:1992 Inc Amendments Nos 1 and 2

Deviations from standard: None

Aim: IP65

ST&C test procedure: LOP 220

ST&C test reports: N58D32683A and 13/0856

Sample delivery date: 12th December 2013

Tests conducted between: 12th December 2013 and 10th January 2014

This certificate refers to the performance of the test sample when tested against the agreed programme. It does not imply that any other samples or products necessarily comply with the requirements of the test programme.

Sira Test & Certification Limited being a UKAS accredited Test House in accordance with ISO/IEC 17025 has tested the sample listed above and found that it met the requirements of the Ingress Protection Code: IP 65

Dated 10th February 2014

S P CORK
Technical Manager

This certificate may only be reproduced in its entirety and without change

Page 1 of 1

Certificate No: Sira 58D32683A
FORM 6240, Issue 8

a CSA Group Company