



CSA Group Testing UK Ltd,  
Unit 6, Hawarden Industrial Park,  
Hawarden, Deeside,  
CH5 3US

**TEST CERTIFICATE  
ISSUED BY CSA GROUP TESTING UK LIMITED**

**TEST FOR THE INGRESS PROTECTION OF AN AreaRAE2 GAS MONITOR  
AND RAEmet WIND SENSOR**

**Manufacturer:** RAE Systems  
3775 North First St  
SAN JOSE, California  
95134  
USA

**Model or Type Identification:** AreaRAE2 Gas monitor and RAEmet Wind sensor

**Standard:** IEC 60529:2013

**Deviations from Standard:** None

**Aim:** IP65

**CSA Test Procedure:** LOP 220

**CSA Test Reports:** N70074642A & 16/0191

**Samples Delivery Dates:** 3<sup>rd</sup> May 2016

**Tests Conducted Between:** 24<sup>th</sup> May and 7<sup>th</sup> June 2016

This certificate refers to the performance of the test samples when tested against the agreed programme. It does not imply that any other samples or products necessarily comply with the requirements of the test programme.

CSA Group Testing UK Limited being a UKAS accredited Test House in accordance with ISO/IEC 17025 has tested the above units and have found them to meet the requirements of the Ingress Protection Code: IP 65

Dated 9<sup>th</sup> June 2016

**S P CORK**  
Technical Manager

*This certificate may only be reproduced in its entirety and without change*

Page 1 of 1

**Certificate No: CSAUK 70074642ALab**  
FORM 6240, Issue 11