

Midas[®] SENSOR CARTRIDGE SPECIFICATIONS

Nitrogen Trifluoride NF₃ and Methyl Fluoride CH₃F Midas-S-XHF, Midas-E-XHF



| Gas Measured | Nitrogen Trifluoride (NF ₃) |
|----------------------------------|--|
| Cartridge Part Number | MIDAS-S-XHF 1 year standard warranty MIDAS-E-XHF 2 year extended warranty |
| Sensor Technology | 3 electrode electrochemical cell |
| Measuring Range (ppm) | NF ₃ 0 - 40 ppm |
| Minimum Alarm 1 Set Point | 4.0ppm |
| Repeatability | < ± 10% of measured value |
| Linearity | < ± 20% of measured value |
| Response Time t _{62.5} | < 110 seconds |
| Sensor Cartridge Life Expectancy | ≥ 24 months under typical application conditions |
| Operating Temperature | 0°C to +40°C (32°F to 104°F) |
| Effect of Temperature | < ± 0.002ppm / °C (at 0°C to 20°C) < ± 0.008ppm / °C (at 20°C to 40°C) |
| Zero Sensitivity | < ± 0.4% of measured value / °C |
| Operating Humidity (continuous) | 20 – 75% rH ¹ |
| Effect of Humidity | Zero < ± 0.003ppm / % rH Sensitivity < ± 1% of measured value / % rH |
| Operating Pressure | 90 - 110kPa |
| Effect of Position | No effect in typical application |
| Long Term Drift | Zero No drift Sensitivity < 15% of measured value / year |
| Calibration Gas | Hydrogen Fluoride (HF) |
| Challenge Gas (Bump Test) | Chlorine (Cl ₂) |
| Warm Up Time | < 20 minutes |
| Storage Temperature | +5°C to +25°C (+41°F to +77°F) |

The sensor data listed is based on ideal test environment; observed performance may vary based on the actual monitoring system and the sampling conditions employed.

Separate Pyrolyzer module (MIDAS-T-NP1) required with the Nitrogen Trifluoride sensor cartridge to detect NF₃ by thermal breakdown. To maintain stated performance, it is recommended to perform gas calibration every 6 months, and ensure the constant temperature of the installation point is in 50 – 104°F(10 - 40°C) and the humidity is in 30 – 70 %RH. Otherwise, more frequent bump testing or calibration will be required to confirm working specifications.

Other Detectable Gases

The following additional gases can be detected with this sensor cartridge. Sensor performance and characteristics will be representative of the data as tabulated above. Consult the Technical Manual to set up the Midas[®] transmitter with the designated identification code for each of the following gas types.

| Detectable Gas | Chemical Formula | Measuring Range |
|-----------------|-------------------|-----------------|
| Methyl Fluoride | CH ₃ F | 0 – 120ppm |

Cross Sensitivities

Each Midas[®] sensor is potentially cross sensitive to other gases and this may cause a gas reading when exposed to other gases than those originally designated. The table below presents typical readings that will be observed when a new sensor cartridge is exposed to the cross sensitive gas (or a mixture of gases containing the cross sensitive species).

| Gas / Vapor | Chemical Formula | Concentration applied (ppm) | Reading (ppm NF ₃) |
|-------------------|----------------------------------|-----------------------------|--------------------------------|
| Arsine | AsH ₃ | 1 | 0 |
| Carbon Monoxide | CO | 2000 | 0 |
| Chlorine | Cl ₂ | 5 | 13.7 |
| Diborane | B ₂ H ₆ | 1 | -1.3 |
| Hydrogen | H ₂ | 20000 | 0 |
| Hydrogen Chloride | HCl | 8 | 14 |
| Hydrogen Fluoride | HF | 2 | 8 |
| Hydrogen Sulphide | H ₂ S | 25 | -3.6 |
| Iso Propanol | C ₃ H ₇ OH | 500 | 0 |
| Methanol | CH ₃ OH | 500 | 0 |
| Nitrogen Dioxide | NO ₂ | 5 | 2.6 |
| Phosphine | PH ₃ | 1 | -0.14 |
| Sulphur Dioxide | SO ₂ | 10 | 22.8 |

Interference differs from cartridge to cartridge and over cell life. It is not recommended to calibrate with cross sensitivity factors. The target gas should be used for calibration.

Find out more

www.honeywellanalytics.com

Toll-free: 800.538.0363

Please Note:

While every effort has been made to ensure accuracy in this publication, no responsibility can be accepted for errors or omissions. Data may change, as well as legislation, and you are strongly advised to obtain copies of the most recently issued regulations, standards, and guidelines. This publication is not intended to form the basis of a contract.